





Occlusal Analysis in Natural Dentition: Systematic Review

Byron Velásquez¹  María Rodríguez²  Verónica Mosquera²  Eddy Álvarez²  Luis Chauca³ 
Alexandra Mena⁴ 

¹ Department of Prosthesis, School of Dentistry, Universidad de Las Americas, Quito, Ecuador

² Department of Prosthesis, School of Dentistry, Universidad de Las Americas, Quito, Ecuador

³ Department of Periodontics and Implantology Oral Research, College of Dentistry, Universidad de Guayaquil, Guayaquil, Ecuador

⁴ Departamento de Investigación de Prótesis Dental, Facultad de Odontología, Universidad de Las Américas, UDLACP, Quito, Ecuador

Address for correspondence Velasquez Ron Byron, PhD, MSc, Department of Prosthesis, School of Dentistry, Universidad de Las Americas (UDLA), Av. Colón y 6. Diciembre CP 170523, Quito, Ecuador (e-mail: byron.velasquez@udla.edu.ec; drvelasquezron@gmail.com).

Eur J Dent 2023;17:615–622.

Abstract

This study aimed to evaluate the effectiveness of conventional occlusal analysis in contrast with digital occlusal analysis in natural dentition. Occlusal analysis allows the identification of normal and abnormal occlusal contact points that alter the cranio-mandibular cervical system. We searched for articles with keywords [[dental occlusion]], [[natural dentition]], [[occlusal adjustment]], [[Immediate Complete Anterior Guidance Development]] [[mastication]], [[bite force]], [[premature contact]], [[occlusal balance]] [[articulating paper]], [[spray]], [[Occlusal contacts]], and [[bite strength]]. They were considered observational, odds ratio and case control studies. We found 189 items. After evaluating the abstracts and full texts of the articles, 10 papers met the inclusion criteria. It was found that occlusal analysis allows the identification of the relationship between poor occlusion and the sensitivity of the teeth due to occlusal trauma, which is also related to temporomandibular joint pain in dynamic occlusion. The contacts of greater strength were observed in nonfunctional cusps, 48%, without ruling out the functional cusps, 24%. Despite being the universal method of occlusal control to date, the use of joint paper, remains subjective compared to the digital occlusal control device. Posture is considered directly related to occlusal trauma and temporomandibular disorders; without proper occlusal analysis, a clear diagnosis of the patient's joint condition cannot be obtained. Digital occlusal analysis is more objective than traditional occlusal analysis.

Keywords

- ▶ dental occlusion
- ▶ occlusal adjustment
- ▶ temporomandibular joint
- ▶ mastication
- ▶ bite force

Introduction

Occlusion is morpho-physiologically static, dynamic, and uniaxial and can be altered by modifications in the stomatognathic system, teeth, muscles, ligaments, and temporomandibular joint.^{1,2} Occlusive analysis allows the identification of normal and abnormal occlusal contact

points. In occlusion, marks are used to indicate the location of contact points after asking the patient to make mandibular movements such as opening, closing, lateral movements (right and left), protrusion, and retrusion.² Marks can be made by spray or paper. The fidelity of these systems depends on the operator's ability, the sensitivity of the marker, etc.^{3,4} When thinner articulating paper is used,

article published online
October 17, 2022

DOI <https://doi.org/10.1055/s-0042-1755626>.
ISSN 1305-7456.

© 2022. The Author(s).

This is an open access article published by Thieme under the terms of the Creative Commons Attribution License, permitting unrestricted use, distribution, and reproduction so long as the original work is properly cited. (<https://creativecommons.org/licenses/by/4.0/>)

Thieme Medical and Scientific Publishers Pvt. Ltd., A-12, 2nd Floor, Sector 2, Noida-201301 UP, India

unnecessary wear on teeth can be prevented. Additionally, some commercial brands do not adhere to the thickness guidelines of occlusive paper.^{5,6} Computerized digital occlusal analysis appears to be a more precise method; it uses technological benefits to achieve the correct occlusal diagnosis, measure relative occlusal forces in static and dynamic mandibular positions, and generate three-dimensional graphics that identify the areas of greater or less pressure.⁷ Despite the advantages of the digital method, articulating paper is the most used method by clinical dentists. By varying occlusal effects in endodontic teeth and utilizing digital or traditional occlusal analysis,⁸ we determined that occlusal contact can be reduced with occlusal adjustment. A similar analysis of occlusal force is measured in overdentures in total edentulous patients, indicating greater force measured with a digital system in the posterior sector.⁹ Simultaneous occlusal contacts observed with articulating paper¹⁰ and digital occlusal analysis¹¹ allow us to differentiate moments of occlusion and disocclusion, which appear visually on the teeth with the help of articulating paper, as well as in the digital occlusal analysis.¹² The aim of this study is to evaluate the effectiveness of conventional occlusal analysis in contrast with digital occlusal analysis in natural dentition.

Methods

Investigators performed a systematic review of selected quantitative studies. The review was performed in accordance with the Preferred Reporting Items for Systematic Review and Meta-Analysis (PRISMA) statement. The review protocol was registered *a priori*, is published online in the International Prospective Register of Systematic

Reviews (PROSPERO) database registration number: CRD42020146370.

The PICO question was as follows: To determine the efficacy of occlusal analysis in natural dentition with joint paper and digital occlusal analysis. Population: Patients with occlusal pathologies. Intervention: Intervention of all patients with occlusal pathologies. Comparison: Comparison of occlusal analysis in natural dentition with joint paper and computerized analysis. Outcome: Observation of mandibular crane cervix after occlusal analysis. [[Observational studies]], [[case-control]], [[odds ratio]], research by keywords [[dental occlusion]], [[natural dentition]], [[Immediate Complete Anterior Guidance Development]], [[occlusal adjustment]], [[mastication]], [[bite force]], [[premature contact]], [[occlusal balance]], [[articulating paper]], [[spray]], [[Occlusal contacts]], [[bite strength]] using the Boolean operators AND, OR, and NOT. Scopus, EBSCO, PubMed, Medline Embase, Cochrane Library, and Web Science and alternate databases Scielo, Latindex, and Redalyc were searched. Preferred Reporting Items for Systematic Review and MetaAnalysis Protocols (PRISMA) research protocol, fluxogram that explains the sequence of information selection.

Inclusion criteria were as follows: complete articles, dates of the articles, odds ratio, case-control with confidence interval (CI) lower and CI upper, and articles containing studies of occlusal pathology. In total, the keyword search returned $n = 189$ articles from the database and additional records identified from other sources were $n = 10$. Some articles were removed after they were found to be duplicate ($n = 115$ records), and some were excluded for not presenting their information on occlusal pathologies, anterior guide, occlusal trauma, or papers that related to bad occlusion with




Fig. 1 Selection Studies flow diagram.

Table 1 Summary of the risk of bias according to Consolidated Standards of Reporting Trials

Item	Arslan et al ⁴ 2017	Basson et al ⁵ 2020	Brizuela-Velasco et al ⁶ 2015	Gupta et al ¹¹ 2019	Kerstein and Radke et al ¹⁴ 2017	Kerstein and Radke et al ¹⁵ 2012	Kerstein and Radke et al ¹⁶ 2014	Kerstein ¹⁷ 2008	Prafulla et al ²⁸ 2020	Yiannios et al ⁴⁸ 2017
1 Abstract	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
2a Background and objectives	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
2b Background and objectives	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
3 Intervention	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
4 Outcomes	Yes	Yes	No	No	Yes	No	Yes	Yes	Yes	Yes
5 Sample size	No	No	Yes	No	Yes	No	Yes	No	Yes	Yes
6 Randomization: sequence generation	No	No	No	No	No	No	No	No	No	No
7 Allocation concealment mechanism	No	No	No	No	No	No	No	No	No	No
8 Implementation	No	No	No	No	No	No	No	No	No	No
9 Blinding	No	No	No	No	No	No	No	No	No	No
10 Statistical methods	Yes	Yes	Yes	No	Yes	Yes	Yes	Yes	Yes	Yes
11 Results: outcomes and estimation	Yes	Yes	No	Yes	Yes	Yes	Yes	Yes	Yes	Yes
12 Discussion: limitations	Yes	No	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
13 Other information: funding	No	No	No	Yes	Yes	Yes	Yes	No	No	No
14 Protocol	Yes	No	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes

Table 2 Summary of the risk of bias

Item	Arslan et al ⁴ 2017	Kerstein and Radke et al ¹⁴ 2017	Prafulla et al ²⁸ 2020	Yiannios et al ⁴⁸ 2017	Gupta et al ¹¹ 2019
Random sequence generation	Low	Low	Low	Low	Unclear
Allocation concealment	Low	Low	Low	Unclear	Unclear
Selective reporting	Low	Low	Low	Low	Low
Blinding (participants and personnel)	High	high	high	High	High
Blinding (outcome assessment)	High	High	High	High	High
Incomplete outcome data	Low	Low	Low	Low	Low

bruxism or temporomandibular disorders. Articles were narrowed down to 84 after screening and excluded 62 of these so that the total number of studies included in the qualitative synthesis was 22. The total number of studies included in the final analysis was 10 (► Fig. 1).

This systematic review was carried out independently by two researchers different from the main one, specialists in the area of oral rehabilitation. Disagreements about study selection were resolved by consensus or intervention of the main researcher. Agreement of the selected articles by the two authors was calculated using kappa coefficient, which was 0.85. In all cases of concordance assessment, the response options were dichotomized to two response options: inclusion and exclusion. As a result, 115 articles were excluded. Of these remaining 84 articles, complete text could be not accessed in 62 articles, only the summaries, and the remaining 22 were completely reviewed by the principal investigator, excluding 25 articles for not meeting the selection criteria due to not adequately describing the clinical criteria of the diagnosis. A final total of 10 articles were obtained.

The authors (B.V.V.R., L.C.H.B., M.R.T., and V.M.C.) independently selected the titles and summaries, excluded duplicates and irrelevant articles, and considered only full-text articles. The date and names of all authors were included in the final review article. Any conflict with respect to inclusion and exclusion criteria was resolved by the third and fourth authors (A.M.S., E.A.). The data extraction procedure was evaluated according to the criteria of all authors. Articles are classified by the author/year, study objective, study type, methodology, the results (standard mean and deviation), and conclusions.

Results

To analyze the research articles, the risk of bias was assessed. Bias is assessed as a judgment for individual elements from domains (selection, reporting, and other) (► Table 1). When assessing to clarify the gold standard, the

collection of information was not homogeneous between the studies, and none reported having considered the reliability and validity of the clinical aspects. The classification system had a dental evaluation to classify patients and was able to apply conventional occlusal analysis or digital analysis. Methods varied between questionnaires (with dichotomous yes/no, multiple-choice, or visual analog scale responses), interviews, the use of immediate complete anterior guidance development, and digital occlusal analysis (► Table 2). Acceptable intra- and interexaminer reliability considered the previous training of examiners prior to clinical measurements; in many of the studies, the number of examiners was unclear and appeared to vary between one, two, or more. The use of adequate samples varied widely in the number of subjects, although a good number of the reports included more than 30 participants. Samples also varied in age ranges, and some of the studies included patients younger than 30 years, while others included patients over 50 years. When assessing replicability of the studies, the studies that presented similar conclusions did not necessarily obey a replicability of the responses of the patients examined in the different studies, which was due to the heterogeneity of the designs and other methodological shortcomings. Without an appropriate occlusal analysis, a clear diagnosis of the patient's joint condition cannot be obtained (► Table 3).

Discussion

Various methods and materials found in the literature that have been used for occlusal analysis have been described.¹³ However, more work is needed in this field by developing new devices or improving them to ensure accurate diagnostic methods that minimize the possibility of error.¹⁴

Qualitative analysis devices are used for clinical detection at the location of the contact point in its exact position¹⁵; this type of material has the advantage of presenting a low cost and simple application, such as static occlusal devices.^{16,17}

Table 3 Structural summary of systematic review

Study (year)	Object of research	Intervention	Evaluation methods	Result
Arslan et al ⁴ 2017	Effect of various occlusal reduction levels	104 dental students	Immediate complete anterior guidance development (ICAGD) coronoplasty	Occlusal reduction ($p > 0.005$)
Basson et al ⁵ 2020	Visually select the 2 highest force occlusal contacts	2 choices as being the highest force 9 different contact areas	T-Scan	12 correct highest force (4.8% correct; 95.2% incorrect). 8 1st choices were correct (6.5% correct; 93.5% incorrect), 4 2nd choices were correct (3.2% correct; 96.7% incorrect)
Brizuela-Velasco et al ⁶ 2015	Occlusal contact surface registered with an articulating paper	15 patients	Articulating paper to obtain their occlusal registrations	Articulating papers (12- μ m or 40- μ m) can avoid unnecessary grinding on teeth during occlusal adjustment
Gupta et al ¹¹ 2019	Compare the occlusal contacts in dentate and edentulous patients	15 dentate and 15 edentulous	T-Scan III	Area was found in dentate and denture wearers, i.e., tooth contact area varies with head posture
Kerstein and Radke et al ¹⁴ 2017	To determine if reducing long disocclusion time	29 muscularly symptomatic	Immediate complete anterior guidance development (ICAGD) coronoplasty	ICAGD time reductions (2.11–0.45 s; $p = 0.0000$) muscle changes ($p = 0.000001$) peak amplitude ($p = 0.00005$) peak contraction ($p < 0.000004$) 50% peak contraction ($p < 0.00001$) silent periods per side (right $p < 0.0000002$ left $p < 0.0000006$) centric occlusion ($p < 0.002$) chewing velocities increased ($p < 0.002$; $p < 0.00005$) opening and closing time: decreased ($p < 0.004$ –0.0001)
Kerstein and Radke et al ¹⁵ 2012	Occlusal adjustment procedure, immediate complete anterior guidance developed (ICAGD)	45 chronic myalgic temporomandibular disorder (TMD) patients.	Medical history and clinical examination. Joint vibration analysis	Disocclusion time to less than 0.4 second ($p < 0.00014$) after Bonferroni correction ($p < 0.0006$)
Kerstein and Radke et al ¹⁶ 2014	Subjective interpretation of paper marking are a reliable method	295 clinicians selected	T-Scan, articulating paper to obtain their occlusal registrations	295 dentists only chose 12.5–13.3% correct contacts by looking at the paper marks ($p < 0.05$)
Kerstein ¹⁷ 2008	Performance of a new design of occlusal sensor, high dentition (HD) sensor	40 maxillary casts with a known arch half force	Computer-guided refinement occlusal adjustments (T-Scan III)	20% reduction in mean jerkiness during opening ($p = 0.0002$), 19% reduction in mean jerk during closing ($p = 0.0114$) were found after ICAGD
Prafulla et al ²⁸ 2020	Occlusal adjustment study using a randomized clinical trial (RCT)	104 first year dental students	Immediate complete anterior guidance development (ICAGD) coronoplasty	1 week retest ($p < 0.05$) Pretreatment through 6 months ($p > 0.05$) 1 week after ICAGD ($p < 0.00001$) 3 months retest ($p < 0.05$)
Yiannios et al ⁴⁸ 2017	After subjects undergo the immediate complete anterior guidance development (ICAGD) computer-guided occlusal adjustment	100 chronically dysfunctional patients	Immediate complete anterior guidance development (ICAGD) coronoplasty spectrophotometer	Discussion time reductions (2.11–0.55 s; $p = 0.0000$)

Table 4 Occlusal indicators

System	Thickness	Localization	Force	Time
Articulating paper	20 µm	+	+	+
T-Scan system	60 µm	+	+	–
Prescale dental	98 µm	+	–	–

With quantitative analysis devices such as digital systems dynamics,^{18,19} the exact intensity has to be registered in the time of contact; however, this system does not locate the position of the contact point, as the sensor is based on standardized estimates of dental positions. However, with the new software updates^{20,21} and the latest digital occlusal system, the individualization of the arcade with greater precision is possible.²² For all these reasons, it is now necessary to use the combination of both systems, qualitative and quantitative^{23,24}, as is the case with the role of articulation in conjunction with digital occlusal systems.

Malta et al compared the sensibility and coincidence of some studies²⁵ between several dispositive qualitative and quantitative studies, concluding that there are differences in the number of contacts between the three systems.²⁶ Both studies found a lower sensitivity of the digital system T-S which can be compared to the other systems.^{27,28} Sensitivity is adversely affected when the sensors are used several times, while the role of articulation is realized at once.^{29,30} In the other systems, it was observed that the number of registered contacts increased when there was no saliva on the occlusal surfaces, while this did not seem to affect the digital occlusal system.^{31,32} One of the limitations of the type of device that uses sensor,³³ refers to when the thickness of the film of the device³⁴ is greater than it should be, the results that are obtained present older contacts of posterior teeth than in previous teeth.^{35,36}

Song et al demonstrated that the appropriate thickness of an occlusive is that of a joint paper to detect previous and subsequent anterior and posterior contacts³⁷ but without altering the patient's proprioception when biting. Thus, it has to be equal to or less than 20 µm.³⁵ A reviewed index assessed the reproducibility and usefulness of instruments that valued occlusion,^{35,38} among them were both static and dynamic devices. It was observed how electromyographic records³⁹ could measure the activation of the jaw musculature and its position, although the cost–benefit relationship of this type of device is not clear.⁴⁰ Finally, the authors concluded that none of the instruments reviewed in the different publications presented results that validated their use and reproducibility.^{41,42}

Some authors recommend that the collection of static and dynamic occlusal data⁴³ did not have reliability by the operator, resulting in problems in the design study⁴⁴ and the lack of standardization.

The most common indicators of static occlusion used by dentists are role of paper articulating,⁴⁵ seda, and nylon, in clinical practice⁴⁶ which allows subjectivity at the moment of occlusal analysis, even if the mark of the indicators is considered and similarity in the variants between the digital systems

(► **Table 4**). Variations in the mark may result in differences among observers (dentists),⁴⁷ while the clinician may remove tooth structure during an occlusal adjustment, another may not see it as convenient. The clinical judgment will depend not only on the knowledge about the occlusion of the individual himself but also on the mark that the indicator used.⁴⁸

The occlusion is not static but has dynamism; consequently, it must be recorded⁴⁹ and the information obtained from the static indicators for the analysis of the occlusal contacts should be complemented with that of dynamic indicators such as digital occlusal systems.^{50,51}

Conclusion

Digital occlusal analysis in most interventions is more objective than traditional occlusal analysis.

Author Contribution

1. Conceptualitation	B.V.R.
2. Data curation	A.M.S./E.A.L.
3. Formal Analysis	B.V.R., V.M.C.
	L.C.B., M.R.T.
4. Acquisition of funds	B.V.R., A.M.S.
5. Research:	B.V.R., V.M.C.,
	L.C.B., M.R.T.
	E.A., A.M.S.
6. Methodology	B.V.R., A.M.S.
7. Project administration	B.V.R.
8. Means	V.M.C., L.C.B., M.R.T., E.A.
9. Software	A.M.S.
10. Supervisión	B.V.R., A.M.S.
11. Validation	B.V.R., V.M.C.
	L.C.B., M.R.T.
	E.A., A.M.S.
12. Visualitation	B.V.R., A.M.S.
13. Drafting – original draft	B.V.R., V.M.C.
	L.C.B., M.R.T.
	E.A., A.M.S.
14. Drafting – review and edition	All the authors

Clinical Relevance

This study tries to remind professionals of the importance of occlusal analysis considering it a mandatory diagnostic means in any multidisciplinary preclinical process, determining that digital occlusal analysis limits the subjectivity of conventional occlusal analysis.

Funding

All funds used to support this work were allocated by the Universidad de Las Americas.

Conflict of Interest

None declared.

Acknowledgment

The authors would like to express their special thanks to Universidad de Las Americas.

Reference

- Abarza L, Sandoval P, Flores M. Digital interocclusal registration in oral rehabilitation using the "T-Scan® III System". A literature review. *Rev Clin Periodoncia Implantol Rehabil Oral* 2016;9(02):95–101
- Afrashtehfar KI, Qadeer S. Computerized occlusal analysis as an alternative occlusal indicator. *Cranio* 2016;34(01):52–57
- Andrus R, Qian F, Schneider R, Huber L, Kerstein R. Comparison of results of traditional occlusal adjustment technique with computer-aided occlusal adjustment technique. *Advanced Dental Technologies & Techniques* 2019;7(01):43–53
- Arslan H, Seckin F, Kurklu D, Karatas E, Yanikoglu N, Capar ID. The effect of various occlusal reduction levels on postoperative pain in teeth with symptomatic apical periodontitis using computerized analysis: a prospective, randomized, double-blind study. *Clin Oral Investig* 2017;21(03):857–863
- Basson E, Kerstein R, Radke R. Ability to correctly select high force occlusal contacts from articulating paper markings. *Advanced Dental Technologies & Techniques* 2020;2(01):101–110
- Brizuela-Velasco A, Álvarez-Arenal Á, Ellakuria-Echevarria J, del Río-Highsmith J, Santamaría-Arrieta G, Martín-Blanco N. Influence of articulating paper thickness on occlusal contacts registration: a preliminary report. *Int J Prosthodont* 2015;28(04):360–362
- Cioffi I, Farella M, Festa P, Martina R, Palla S, Michelotti A. Short-term sensorimotor effects of experimental occlusal interferences on the wake-time masseter muscle activity of females with masticatory muscle pain. *J Oral Facial Pain Headache* 2015;29(04):331–339
- Ferreira FM, César Simamoto-Júnior P, Soares CJ, Ramos AMAM, Fernandes-Neto AJ. Effect of occlusal splints on the stress distribution on the temporomandibular joint disc. *Braz Dent J* 2017;28(03):324–329
- Ficnar T, Middelberg C, Rademacher B, Hessling S, Koch R, Figgenger L. Evaluation of the effectiveness of a semi-finished occlusal appliance—a randomized, controlled clinical trial. *Head Face Med* 2013;9(01):5
- Nishimori H, Iida T, Kamiyama H, et al. Effect of sleep restriction on somatosensory sensitivity including occlusal sensation in the orofacial area. *J Prosthodont Res* 2019;63(02):193–198
- Gupta S, Tarannum F, Gupta N, Upadhyay M, Abdullah A. Effect of head posture on tooth contacts in dentate and complete denture wearers using computerized occlusal analysis. *J Indian Prosthodont Soc* 2019;19(01):79–82
- Heredia-Rizo AM, Oliva-Pascual-Vaca A, Rodríguez-Blanco C, Piña-Pozo F, Luque-Carrasco A, Herrera-Monge P. Immediate changes in masticatory mechanosensitivity, mouth opening, and head posture after myofascial techniques in pain-free healthy participants: a randomized controlled trial. *J Manipulative Physiol Ther* 2013;36(05):310–318
- Paesini D, Guarda/Nardini C, Salmasso L, Manfredini D. Reliability of multiple-degree incisal/occlusal tooth wear assessment on dental casts: findings from a five examiner investigation and related clinical implications. *J Oral Rehabil* 2017;31(03):242–250
- Kerstein RB, Radke J. Average chewing pattern improvements following disclusion time reduction. *Cranio* 2017;35(03):135–151
- Kerstein RB, Radke J. Masseter and temporalis excursive hyperactivity decreased by measured anterior guidance development. *Cranio* 2012;30(04):243–254
- Kerstein RB, Radke J. Clinician accuracy when subjectively interpreting articulating paper markings. *Cranio* 2014;32(01):13–23
- Kerstein RB. T-scan III applications in mixed arch and complete arch, implant -supported prosthodontics. *Dent Implantol Update* 2008;19(07):49–53
- Khuder T, Yunus N, Sulaiman E, Ibrahim N, Khalid T, Masood M. Association between occlusal force distribution in implant overdenture prostheses and residual ridge resorption. *J Oral Rehabil* 2017;44(05):398–404
- Koh H, Robinson PG. Occlusal adjustment for treating and preventing temporomandibular joint disorders. *Cochrane Database Syst Rev* 2003;1(01):CD003812
- Koos B, Godt A, Schille C, Göz G. Precision of an instrumentation-based method of analyzing occlusion and its resulting distribution of forces in the dental arch. *J Orofac Orthop* 2010;71(06):403–410
- Koos B, Höller J, Schille C, Godt A. Time-dependent analysis and representation of force distribution and occlusion contact in the masticatory cycle. *J Orofac Orthop* 2012;73(03):204–214
- Kürklüarpacay D, Bayindir F, Diñçkalyanikoğlu N. Evaluation of premature contacts using the T-Scan III. *Open J Stomatol* 2019;9:136–145
- Lin PT, Jiao Y, Zhao SJ, et al. Occlusion and disocclusion time changes in single unit crowns designed by functional generated path technique: a randomised clinical trial. *Sci Rep* 2017;7(01):388
- Lino P, Martins C, Miranda G, et al. Reliability of multiple-degree incisal/occlusal tooth wear assessment on dental casts: findings from a fiveexaminer investigation and related clinical implications. *J Oral Rehabil* 2018;45(03):539–544
- Malta Barbosa J, Urtula AB, Hirata R, Caramês J. Thickness evaluation of articulating papers and foils. *J Esthet Restor Dent* 2018;30(01):70–72
- Manfredini D, Vano M, Peretta R, Guarda-Nardini L. Jaw clenching effects in relation to two extreme occlusal features: patterns of diagnoses in a TMD patient population. *Cranio* 2014;32(01):45–50
- Mitchem JA, Katona TR, Moser EAS. Does the presence of an occlusal indicator product affect the contact forces between full dentitions? *J Oral Rehabil* 2017;44(10):791–799
- Prafulla T, Shwetha P, Bhavya B, Mounika A, Kerstein R, Radke J. A disclusion time reduction randomized controlled occlusal adjustment trial. *Advanced Dental Technologies & Techniques* 2020;2(02):2–22
- Kuzmanovic Pfcifer J, Dodic S, Lasic V, Trajkovic G, Milic N, Milicic B. Occlusal stabilization splint for patients with temporomandibular disorders: Meta-analysis of short and long term effects. *PLoS One* 2017;12(02):e0171296
- Qadeer S, Kerstein R, Kim RJ, Huh JB, Shin SW. Relationship between articulation paper mark size and percentage of force measured with computerized occlusal analysis. *J Adv Prosthodont* 2012;4(01):7–12
- Qadeer S, Abbas A, Sarinnaphakorn L. Comparison of excursive occlusal force parameters in post orthodontic and non orthodontic subjects using T Scan III.pdf. *Cranio* 2016;34(06):395–401
- Da Costa FF, Santos GSMQ, Farias-Neto A, Sanchez-Ayala A, Rizzatti-Barbosa CM. The relationship between occlusal support and maxillary development: an animal study. *Eur J Dent* 2015;9(03):400–403
- Sabir S, Regragui A, Merzouk N. Maintaining occlusal stability by selecting the most appropriate occlusal scheme in complete removable prosthesis. *Jpn Dent Sci Rev* 2019;55(01):145–150

- 34 Skármeta NP. Occlusal stability and mandibular stability: the major part of dentistry we are still neglecting. *Cranio* 2017;35(04):201–203
- 35 Solow RA. Clinical protocol for occlusal adjustment: rationale and application. *Cranio* 2018;36(03):195–206
- 36 Sierpinska T, Golebiewska M, Lapuc M. The effect of mastication on occlusal parameters in healthy volunteers. *Adv Med Sci* 2008;53(02):316–320
- 37 Song JH, Cho ES, Kim ST, Ahn HJ. Change of distribution and timing of bite force after botulinum toxin type A injection evaluated by a computerized occlusion analysis system. *Yonsei Med J* 2014;55(04):1123–1129
- 38 Sutter B, Yiannios N, Radke J. Letter to the Editor regarding “In vitro evaluation of T-Scan® III through study of the sensels” from CRANIO Volume 33 Issue 4, by Manuel Cerna et al. *Cranio* 2016;34(03):208–209
- 39 Sutter BA. A digital poll of dentists testing the accuracy of paper mark subjective interpretation. *Cranio* 2018;36(06):396–403
- 40 Sutter BA, Radke J. Letter to the Editor regarding “Does the presence of an occlusal indicator product affect the contact forces between full dentitions?” by Mitchem, Katona and Moser *J Oral Rehabil* 2018;45(07):571–573
- 41 Tamaki K, Ishigaki S, Ogawa T, et al. Japan Prosthodontic Society position paper on “occlusal discomfort syndrome”. *J Prosthodont Res* 2016;60(03):156–166
- 42 Tang Y, Li H, Chen Y, Zhu L, Kang H. Effect of different splint thicknesses on occlusal function and temporomandibular joint sounds: a clinical report. *Open J Stomatol* 2018;8:326–337
- 43 Thumati P, Manwani R, Mahantshetty M. The effect of reduced disclusion time in the treatment of myofascial pain dysfunction syndrome using immediate complete anterior guidance development protocol monitored by digital analysis of occlusion. *Cranio* 2014;32(04):289–299
- 44 Chaikla K, Pumklin J, Piyapattamin T. Comparison of occlusal parameters between open bite and nonopen bite patients using the T-Scan III system: a pilot study. *Eur J Dent* 2022;17(01):100–103
- 45 Throckmorton GS, Rasmussen J, Caloss R. Calibration of T-Scan sensors for recording bite forces in denture patients. *J Oral Rehabil* 2009;36(09):636–643
- 46 Ritthiti A, Sattabanasuk V, Karunratanakul K, Senawongse P. Effect of stress generated by occlusal cyclic force on class I bulk-fill composite restoration microleakage. *Eur J Dent* 2021;17(01):100–103
- 47 Wang S, Li Z, Wang G. Preliminary clinical application of complete digital workflows. *Peking Univ Health Sci* 2019;51(01):105–109
- 48 Yiannios N, Kerstein RB, Radke J. Treatment of frictional dental hypersensitivity (FDH) with computer-guided occlusal adjustments. *Cranio* 2017;35(06):347–357
- 49 Ruttitivapanich N, Tansalarak R, Palasuk J, Pumklin J. Correlation of bite force interpretation in maximal intercuspal position among patient, clinician, and T-Scan III system. *Eur J Dent* 2019;13(03):330–334
- 50 Zana D, et al. Differences between centric relation and maximum intercuspation as possible cause for development of temporomandibular disorder analyzed with T-scan III. *Eur J Dent* 2015;9(04):83–289
- 51 Zhu L, Li H, Chen Y, Kang H, Tang Y. Effect of different splint thicknesses on occlusal function and temporomandibular joint sounds: a clinical report. *BioMed Res Int* 2018;(12):326–337