

Over-the-scope clip

Bhanwar Singh Dhandhu, Kumar Shwetanshu Narayan, Surinder Sultania, Sandeep Nijhawan

Department of Gastroenterology, SMS Medical College and Hospital, Jaipur, Rajasthan, India

Abstract

The endoscopic clips were first described by Hayashi in Japan in 1975. Since then, many different types of clips have been introduced to the gastroenterology practice. Over-the-scope clip (OTSC) or the “beer claw” was introduced in the year 2010. It works on the principle of dynamic compression and is applied over the scope like a band ligator. OTSC is a new endoscopic modality which can be used in acute gastrointestinal hemorrhage, iatrogenic perforation, anastomotic leak, and chronic fistula treatment with fairly good success rate and therefore, is now being considered as an alternative to surgery, especially in situations where surgery is not feasible.

Key words

Beer claw and dynamic compression, endoscopic clips, over the scope clip

Introduction

The endoscopic clips were first described by Hayashi in Japan in 1975.^[1] After improvement in the design, it was used by Soehendra in patients of upper gastrointestinal (GI) bleeding for hemostasis.^[2] Later on, these clips were improved as reloadable, preloaded quick clip, and rotatable clip. A new type of clip was introduced in the year 2010, working on a different principle as compared to the previous clips which were through-the-scope (TTS); this clip was over the scope before deployment loaded on a cylinder. The deployment was similar to band ligation, was named as over-the-scope clip (OTSC), and was commonly known as the “bear claw.”

Over-the-scope Clip Device

The device consists of a preloaded cylinder with the clip and thread, a thread retriever, and the release hand wheel [Figure 1].^[3] The clip is made up of nitinol alloy, which allows a high grade of elasticity^[4] and comes in three


different types: Type “a” (rounded) is with blunt spikes and meant for compression effect particularly in the thinner walled esophagus and colon, Type “t” (pointed) has teeth with small spikes which not only compresses the tissue but also has better anchoring, and the third Type “gc” (longer pointed) has elongated teeth with spikes, which is specific for closure of the stomach wall which has maximum thickness as compared to duodenum and colon.

The applicator cap, with a mounted nitinol clip, is affixed to the tip of the endoscope in a manner similar to that of a variceal band ligation cap. The clip fits onto the cylindrical cap in the open position. Caps are available in three diameters to accommodate various endoscope diameters: (1) 11 mm (designed for endoscope diameters 8.5–11 mm), (2) 12 mm (for endoscope diameters 10.5–12 mm), and (3) 14 mm (for endoscope diameters 11.5–14 mm). With the applicator cap attached to the endoscope, the corresponding outer diameter of the instrument is 16.5, 17.5, or 21 mm, respectively. Caps are also available in two depths (3 and 6 mm) to allow variation in the amount of tissue grasped during approximation. The commonly used depth is of 6 mm which has adequate depth for retrieval of the tissue in the cylinder for better deployment of the clip.

Address for correspondence:

Dr. Sandeep Nijhawan, Department of Gastroenterology,
SMS Medical College and Hospital, Jaipur, Rajasthan, India.
E-mail: dr_nijhawan@yahoo.com

Access this article online

Website: www.jdeonline.in	Quick Response Code 
DOI: 10.4103/0976-5042.189127	

This is an open access article distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 3.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as the author is credited and the new creations are licensed under the identical terms.

For reprints contact: reprints@medknow.com

How to cite this article: Dhandhu BS, Narayan KS, Sultania S, Nijhawan S. Over-the-scope clip. J Dig Endosc 2016;7:47-50.

Assist devices

There are three additional instruments available that facilitate the use of OTSC.

Twin grasper

This device has two separate, lateral, independently controlled, mobile jaws [Figure 1]^[3] which help in approximating the wound edges before deploying the clip. It is particularly useful for larger defects and chronic defects associated with indurated tissue where suction alone may be inadequate to approximate tissue. It is available in flexible catheter length of 165 and 220 cm for use with gastroscope and colonoscope, respectively. An endoscope working channel of at least 3.2 mm is required to use this device with the OTSC system. The approximated tissue is then pulled into the applicator cap; additional suction is applied if necessary, and the clip is then deployed by turning the hand wheel.

Anchor

This device is used to retract hard and fibrotic tissue into the cap when simple suction or the twin grasper may not be effective. This accessory consists of a 165 cm long flexible catheter with three retractable needle pins [Figure 1].^[3] On releasing, the three pins pierce tissue along a curved path, thereby anchoring it. The tissue can then be pulled into the applicator cap, suction is applied if necessary, and the clip is deployed. The anchor is then retracted and removed. An endoscope working channel of at least 3.2 mm is required to use this device with the OTSC system.

Reloader

After deployment of the initial clip, this device [Figure 1]^[3] assists in mounting additional clips onto the applicator cap. This device is useful if more than 1 clip is to be placed in the same treatment session.

Functional principle

The OTSC works on the principle of dynamic compression that it compresses the tissue in between its teeth and the space between the teeth allows microperfusion of the clipped tissue. Therefore, there is no necrosis of the clipped tissue. The clip

enables a dynamic closure of the tissue avoiding inadequate closure overtime which usually occurs in fixed suture loops.^[4,5]

Procedure

The clip cylinder is fixed on the tip of the endoscope until the stopper in the cap, if it is not done up to the stopper, it could slide forward when turning the hand wheel which may lead to clipping of the instrument. In case there is failure to do so, the cap should be reapplied till the stopper in the cap under the endoscopic view.

First, the bleeding source should be identified by diagnostic endoscopy, followed by application of the cap with the clip over the tip of the endoscope. The thread is retrieved by the retriever through the suction channel and fixed to the release handle wheel. The bleeding site is identified and sucked into the cap, followed by release of the clip by rotating the release wheel. In patients who develop iatrogenic perforation of the bowel, twin grasper is used to grasp the two edges and the retraction of the tissue in the cap is followed by release of clip and withdrawal of the twin grasper. Patients of chronic inflammatory fistula have fibrosis along with inflammation making it difficult to suck or grasp; therefore, the anchor is required, which not only anchor the tissue but also guide the cap to the lesion which is then retrieved into the cap followed by release of the clip and withdrawal of the anchor.

Discussion

OTSC is a new endoscopic modality which has proven to be safe and effective treatment in patients with GI bleed, perforations, anastomotic leaks, and chronic GI fistulae. In patients presenting with GI bleed who have relapsed after conventional treatment, OTSC has been found to be technically successful in 100% and clinically successful in 94–100% cases [Table 1].^[6-11,21] In patients with iatrogenic perforation and anastomotic leaks, overall technical success of OTSC is 85–100% and clinical success is 80–100% [Table 2].^[6-9,12-18,21] Patients with chronic fistula have fibrosis with chronic inflammation. The technical success in such patients is 75–100% and clinical success is 75–100% [Table 3].^[6,7,12,14-16,21]

With respect to Indian literature on OTSCs, a recent prospective trial by Goenka *et al.* has shown a technical and



Figure 1: The OTSC system^[3]

Table 1: Studies showing success rates of over-the-scope clip in gastrointestinal hemorrhage

Author	Number of lesions	Technical success (%)	Clinical success (%)
Albert <i>et al.</i> ^[6]	7	100 (7/7)	57 (4/7)
Baron <i>et al.</i> ^[7]	7	100 (7/7)	100 (7/7)
Kirschniak <i>et al.</i> ^[8]	7	100 (7/7)	100 (7/7)
Kirschniak <i>et al.</i> ^[9]	27	100 (27/27)	100 (27/27)
Repici <i>et al.</i> ^[11]	7	100 (7/7)	100 (7/7)
Manta <i>et al.</i> ^[10]	30	100 (30/30)	97 (27/30)
Goenka <i>et al.</i> ^[21]	6	100 (6/6)	100 (6/6)

Table 2: Studies showing success rates of over-the-scope clip in gastrointestinal perforation and acute anastomotic leak

Author	Number of lesions	Technical success (%)	Clinical success (%)
Albert <i>et al.</i> ^[6]	5	80 (4/5)	80 (4/5)
Arezzo <i>et al.</i> ^[12]	8	87 (7/8)	87 (7/8)
Baron <i>et al.</i> ^[7]	8	NA	62 (5/8)
Gubler and Bauerfeind ^[13]	14	93 (13/14)	78 (11/14)
Kirschniak <i>et al.</i> ^[8]	4	100 (4/4)	100 (4/4)
Kirschniak <i>et al.</i> ^[9]	11	100 (11/11)	100 (11/11)
Manta <i>et al.</i> ^[14]	5	100 (5/5)	80 (4/5)
Parodi <i>et al.</i> ^[15]	6	88 (5/6)	83 (5/6)
Sandman <i>et al.</i> ^[16]	7	NA	86 (6/7)
Voermans <i>et al.</i> ^[18]	36	92 (33/36)	89 (32/36)
Seebach <i>et al.</i> ^[17]	7	85 (6/7)	57 (4/7)
Goenka <i>et al.</i> ^[21]	3	100 (3/3)	100 (3/3)

NA=Not available

Table 3: Studies showing success rates of over-the-scope clip in chronic gastrointestinal lesions (fistulae)

Author	Number of lesions	Technical success (%)	Clinical success (%)
Albert <i>et al.</i> ^[6]	7	100 (7/7)	71 (5/7)
Arezzo <i>et al.</i> ^[12]	6	83 (5/6)	83 (5/6)
Baron <i>et al.</i> ^[7]	28	NA	71 (20/28)
Manta <i>et al.</i> ^[14]	9	88 (8/9)	88 (8/9)
Parodi <i>et al.</i> ^[15]	4	100 (4/4)	100 (4/4)
Sandman <i>et al.</i> ^[16]	6	NA	83 (5/6)
Goenka <i>et al.</i> ^[21]	4	75 (3/4)	75 (3/4)

NA=Not available

clinical success rate of 100% and 83.3% in GI hemorrhage and GI defects, respectively. This study also highlighted the fact that tattooing to mark the leak site can help in accurate deployment and thus increase the success rate of OTSCs.^[21]

OTSC comprises newer generation clips which have several advantages with respect to TTS clips. Due to its large size and the use of the cap and tissue graspers (forceps or anchor), the OTSC system allows for the entrapment of a larger amount of tissue and provides high stability and closure and minimal strain on surrounding tissue, thereby preventing ischemia or cutting the tissue. The OTSCs is designed to create a full-thickness closure of perforations, whereas a TTS clip can compress only a limited amount of tissue. Accordingly, the hemostatic effect of OTSC is more in comparison with TTS clips. The clinical success also appears to be better than TTS clips in the closure of iatrogenic perforations, and the average number of OTSC clip required is much less than the TTS clips.^[18] While TTS clips are effective for leaks smaller than 10 mm, OTSC are preferred for larger defects (20–30 mm).^[19] However, if reloading of clips is required, OTSC require removal of endoscope whereas TTS clips do not. Another disadvantage is that OTSC are of large diameter and therefore utmost care has to be taken while introducing them through physiologically narrow spaces such as cricopharynx or the

pylorus. There are case reports of iatrogenic perforations with OTSCs which are rare with TTS clip.^[18]

As far as limitations of OTSC are concerned, studies have shown that OTSCs are not suitable for the closure of chronic lesions with hard, severely fibrotic wounds because it is difficult to draw such a lesion into the top of the device. In addition, the OTSCs are also not recommended for defects larger than 30 mm.^[20]

Conclusion

The OTSC clip is a new endoscopic treatment for acute GI hemorrhage, iatrogenic perforation, anastomotic leak, and chronic fistula with fair technical and clinical success rate and may be a good alternative to surgery, especially in situations where surgery is not feasible.

Financial support and sponsorship

Nil.

Conflicts of interest

There are no conflicts of interest.

References

- Hayashi I, Yonezawa TM, Kuwabara T, Kudoh I. The study on staunch clip for the treatment by endoscopy. *Gastrointest Endosc* 1975;17:92-101.
- Binmoeller KE, Thonke F, Soehendra N. Endoscopic hemoclip treatment for gastrointestinal bleeding. *Endoscopy* 1993;25:167-70.
- Kouklakis G, Zazos P, Liratzopoulos N, Gatopoulou A, Oikonomou A, Pitiakoudis M, *et al.* Endoscopic treatment of a gastrocutaneous fistula using the over-the-scope-clip system: A case report. *Diagn Ther Endosc* 2011;2011:384143.
- von Renteln D, Denzer UW, Schachschal G, Anders M, Groth S, Rösch T. Endoscopic closure of GI fistulae by using an over-the-scope clip (with videos). *Gastrointest Endosc* 2010;72:1289-96.
- Raju GS, Gajula L. Endoclips for GI endoscopy. *Gastrointest Endosc* 2004;59:267-79.
- Albert JG, Friedrich-Rust M, Woeste G, Strey C, Bechstein WO, Zeuzem S, *et al.* Benefit of a clipping device in use in intestinal bleeding and intestinal leakage. *Gastrointest Endosc* 2011;74:389-97.
- Baron TH, Song LM, Ross A, Tokar JL, Irani S, Kozarek RA. Use of an over-the-scope clipping device: Multicenter retrospective results of the first U.S. experience (with videos). *Gastrointest Endosc* 2012;76:202-8.
- Kirschniak A, Kratt T, Stüker D, Braun A, Schurr MO, Königsrainer A. A new endoscopic over-the-scope clip system for treatment of lesions and bleeding in the GI tract: First clinical experiences. *Gastrointest Endosc* 2007;66:162-7.
- Kirschniak A, Subotova N, Zieker D, Königsrainer A, Kratt T. The over-the-scope clip (OTSC) for the treatment of gastrointestinal bleeding, perforations, and fistulas. *Surg Endosc* 2011;25:2901-5.
- Manta R, Galloro G, Mangiavillano B, Conigliaro R, Pasquale L, Arezzo A, *et al.* Over-the-scope clip (OTSC) represents an effective endoscopic treatment for acute GI bleeding after failure of conventional techniques. *Surg Endosc* 2013;27:3162-4.
- Repici A, Arezzo A, De Caro G, Morino M, Pagano N, Rando G, *et al.* Clinical experience with a new endoscopic over-the-scope clip system for use in the GI tract. *Dig Liver Dis* 2009;41:406-10.
- Arezzo A, Verra M, Reddavid R, Cravero F, Bonino MA, Morino M. Efficacy of the over-the-scope clip (OTSC) for treatment of colorectal postsurgical leaks and fistulas. *Surg Endosc* 2012;26:3330-3.

13. Gubler C, Bauerfeind P. Endoscopic closure of iatrogenic gastrointestinal tract perforations with the over-the-scope clip. *Digestion* 2012;85:302-7.
14. Manta R, Manno M, Bertani H, Barbera C, Pigò F, Mirante V, *et al.* Endoscopic treatment of gastrointestinal fistulas using an over-the-scope clip (OTSC) device: Case series from a tertiary referral center. *Endoscopy* 2011;43:545-8.
15. Parodi A, Repici A, Pedroni A, Bianchi S, Conio M. Endoscopic management of GI perforations with a new over-the-scope clip device (with videos). *Gastrointest Endosc* 2010;72:881-6.
16. Sandmann M, Heike M, Faehndrich M. Application of the OTSC system for the closure of fistulas, anastomosal leakages and perforations within the gastrointestinal tract. *Z Gastroenterol* 2011;49:981-5.
17. Seebach L, Bauerfeind P, Gubler C. "Sparing the surgeon": Clinical experience with over-the-scope clips for gastrointestinal perforation. *Endoscopy* 2010;42:1108-11.
18. Voermans RP, Le Moine O, von Renteln D, Ponchon T, Giovannini M, Bruno M, *et al.* Efficacy of endoscopic closure of acute perforations of the gastrointestinal tract. *Clin Gastroenterol Hepatol* 2012;10:603-8.
19. Qadeer MA, Dumot JA, Vargo JJ, Lopez AR, Rice TW. Endoscopic clips for closing esophageal perforations: Case report and pooled analysis. *Gastrointest Endosc* 2007;66:605-11.
20. Nishiyama N, Mori H, Kobara H, Rafiq K, Fujihara S, Kobayashi M, *et al.* Efficacy and safety of over-the-scope clip: Including complications after endoscopic submucosal dissection. *World J Gastroenterol* 2013;19:2752-60.
21. Goenka MK, Rai VK, Goenka U, Tiwary IK. Endoscopic management of gastrointestinal leaks and bleeding with the over-the-scope clip: A prospective study. *Clin Endosc* 2016. [Epub ahead of print].